

Technical Specification and Terms & Conditions

Manual Abel Flash Point Apparatus :

- | | |
|----------------------------|---|
| 1. Type | : Bench Top Type |
| 2. Model | : Bidder must mention complete model number |
| 3. Manufacturer | : Bidder must mention manufacturer |
| 4. Country of Origin | : USA, JAPAN, UK, EU |
| 5. Conformity Standard | : IP 170 |
| 6. Test Temperature Range | : -30 °C to 70 °C |
| 7. Temperature Measurement | : PT-100 |
| 8. Resolution | : 0.1 °C |
| 9. Heating | : Electrically by electronic regulator |
| 10. Stirring rate | : Automatic as per method |
| 11. Igniter | : Gas |
| 12. Cooling | : Liquid-Cooled |
| 13. Detection | : Visual |
| 14. Display | : Digital |
| 15. Safety | : Overheat Protection, automatic shut-off |
| 16. User Calibration | : Bidder should specify |
| 17. Power Cable | : With three pin plug |
| 18. Power | : 220 V±10%, 50 Hz |

Terms & Condition :

1. Warranty: One year from the date of delivery.
2. Bidder must submit Manufacturer's product catalogue/brochure /technical data sheet of the offered item
3. Manufacturer Authorization letter is required.
4. Bidder must provide Instruction Manual & Service Manual in English at the time of delivery.
5. At least 10 (ten) years conformation of supplying necessary spare parts and consumables shall be given by the supplier
6. Bidder should provide spare parts list with price.

Services

Installation & Commissioning

